



## SPEEDY NEEDLESS



*... A simple, practical and safe gesture ...*



## **SPEEDY NEEDLESS**

### **PARTICLE FILTRATION SYSTEM IN VIAL OR AMPOULE WITHDRAWAL**



Simple, practical, safe and universal, Speedy Needless is a rational response to particle filtration during locoregional anesthesia in **SINGLE SHOT**. The system allows to filter the anesthetic or any other injectable solution during the aspiration phase without any type of metallic needle. Thanks to a special **0.45µ FILTER**, placed at the end of the luer cone of the universal ampoule, the solution is free from any type of impurity or fragment of glass, plastic or rubber, before the pharmacological bolus is carried out.

Speedy Needless makes filtration easy in any situation where there is no possibility to insert an in-line filter such as **SPINAL** anaesthesia.

Speedy Needless is available individually or with syringes of various sizes in order to reduce connections and disconnections during the aspiration phase, while the special flexible probe makes it easy to attach to any type of glass or plastic vial or ampoule.



Speedy Needless is particularly indicated also during antibiotic therapy and its preparation. The device has been especially designed for aspiration from morphine vials intended for loading elastomeric pumps in order to reduce possible breakage of the same by fragments and to eliminate the risk of accidental punctures during filling operations.

- Multi-purpose ampoule
- Flexible probe with rigid tip
- No metal needles
- Syringe already connected
- Wide range
- Safe and effective
- Universal
- Luer Lock connections
- Transparent
- 5µ Filter

#### **REFERENCES**

Sabon RL Jr, Cheng EY, Stommel KA, Hennen CR. Glass particle contamination: influence of aspiration methods and ampule types. *Anesthesiology*. 1989; 70: 859-62. Department of Anesthesiology, Medical College of Wisconsin, Milwaukee.

Bikhazi AB, Shiatis JA, Haddad AF. Quantitative estimation of particulate matter in pharmaceutical preparations intended for intravenous administration. *J Pharm Sci*. 1977; 66: 181-6.

Walpot H, Franke RP, Burchard WG, et Al. [Particulate contamination of infusion solutions and drug additives within the scope of long-term intensive therapy. I. Energy dispersion electron images in the scanning electron microscope-REM / EDX - (Article in German)] *Anesthesist*. 1989; 38: 544-8. Klinik für Anaesthesiologie Rheinisch-Westfälische Technische Hochschule - Aachen - Germany.

Rinaldi I, et al. Contamination of Pantopaque by glass. *Surg Neurol* 11: 295-7; 1979

Pinnock CA. Particulate contamination of solutions for intrathecal use. *Ann R Coll Surg Engl*. 66: 423; 1984

Cooley RL, et al. Particulate contamination of local anesthetic solutions. *Oral Surg Oral Med Oral Pathol* 51: 481-3; 1981

Eriksen S. Particulate contamination in spinal analgesia. *Acta Anaesthesiol Scand* 32: 545-8; 1988

**Product Code: 2000799**

**Box of 50 Pcs.**